

Subject:	LABORATORY CONDITIONS
Proposal:	Technical Committee 26.06.08 The requirement to carry out the mechanical tests under laboratory conditions (air-conditioned labs) as prescribed in ISO standards only applies to QUALICOAT approved testing laboratories, but not to the licensed plants.
QUALICOAT resolution:	TC Meeting held on 26 June 2008 <i>(ratified on 19 November 2008)</i>
Date of application:	January 2009
Amendments to the Specifications:	Sentence added at the end of paragraph 3.6
3.6 Laboratory [...] Each piece of apparatus must have a data sheet showing the apparatus identification number and calibration checks. The laboratory conditions may differ from those prescribed by ISO standards for mechanical tests.	